

Information

always your partner

EDI™ COVID-19 Nucleocapsid IgG IVD CE Quantitative ELISA Kit

ELISA test for the quantitative determination of COVID-19 Nucleocapsid IgG in serum samples

- COVID-19 IgG concentration instead of OD values;
- Possibility to:
 - Standardize COVID-19 measurement;
 - Compare IgG concentrations during the disease course;
 - Effectively/Accurately monitor the immune response;
- Easier qualification of convalescent plasma donors;

Test principle	Determination of IgG against COVID-19 in patient samples
Technology	Indirect quantitative microplate based immunoassay using 5 calibrator levels
Sample type	Serum
Incubation time	80 minutes
Positive Predicted Value (PPV)	90,3%
Negative Predicted Value (NPV)	97,3%
Specificity	98,3%
Description	EDI™ COVID-19 Nucleocapsid IgG Quantitative ELISA
Produced by	Epitope Diagnostics, Inc, San Diego, USA
Packaging	1 Microtiter plate with 96 wells
Reference	KT-1034

Für weitere Informationen wenden Sie sich bitte an:
For further information please contact:
Pour plus d'information, veuillez contacter:

www.tecomedical.com



A EUROBIO SCIENTIFIC COMPANY

Switzerland / Headquarters

TECOmedical AG

Gewerbestrasse 10

4450 Sissach

Phone +41 61 985 81 00

Fax +41 61 985 81 09

Mail info@tecomedical.com

Germany

TECOmedical GmbH

Wasserbreite 57

32257 Bünde

Phone +49 52 23 985 99 99

Fax +49 52 23 985 99 98

Mail info@tecomedical.com

Benelux

TECOmedical Benelux BV

Computerweg 22

3542 DR Utrecht, The Netherlands

Phone +31 30 307 87 30

Fax +31 30 307 49 39

Mail benelux@tecomedical.com

Austria

TECOmedical AG

Phone 0800 20 40 66

Fax 0800 20 40 55

Mail info@tecomedical.com